

AEC Chromogen Concentrate

Description:

3-Amino-9-Ethylcarbazole, AEC Chromogen Concentrate is a widely used chromogen for immunohistochemical staining and immunoblotting. When in the presence of the peroxidase enzyme, AEC produces a vivid red end product that is soluble in alcohol. Stained slides must be mounted in any commercially available mounting media. This product is available in a convenient two component form: AEC Chromogen Concentrate and AEC Substrate. For maximum flexibility, this product is available in bulk amounts or dropper-top. Use with ScyTek AEC Substrate (ACE500 and ACE999).


Uses/Limitations:

Not to be taken internally.
For In-Vitro Diagnostic use only.
Histological applications.
Do not use if reagents become cloudy.
Do not use past expiration date.
Use caution when handling reagents.
Non-Sterile.

Control Tissue:

Any well fixed paraffin embedded or frozen tissue section.

Ordering information regarding individual components on back page!



Kit Options:

Additional Kit Reagents Sold Separately

<u>Item #</u>	<u>Components</u>	<u>Volume</u>	<u>Storage</u>
ACJ500	AEC Chromogen Concentrate	15 ml	2-8°C
	AEC Substrate Buffer	500 ml	2-8°C
ACJ999	AEC Chromogen Concentrate	30 ml	2-8°C
	AEC Substrate Buffer	1000 ml	2-8°C

Precautions:

Avoid contact with skin and eyes.
Harmful if swallowed.
Follow all Federal, State, and local regulations regarding disposal.

Storage: 2° C  8° C

 ScyTek Laboratories, Inc.
205 South 600 West
Logan, UT 84321
U.S.A.

CE 

EC REP
Emergo Europe
Westervoortsedijk 60
6827 AT Arnhem, The Netherlands

Procedure:

- Combine 20µl of AEC Chromogen Concentrate to each 1ml of AEC Substrate Buffer. Combined mixture must be used immediately.

NOTE: Range = 20-30µl of AEC Chromogen / 1 ml of AEC Substrate.

20µl /ml produces optimal staining quality, 30µl/ml produces maximum sensitivity.

- Dip slide one time in DI/Distilled water immediately prior to application of AEC.
- Apply mixture to tissue section.
- Incubate tissue section for 5 minutes.
- Rinse slide using DI/Distilled water.
- Apply mixture and incubate for another 5 minutes.
- Rinse slide in two changes of DI/Distilled water.
- If needed, repeat application, incubation, and rinse for a third 5 minute period.
- Counterstain as desired.
- Rinse slide in two changes of DI/Distilled water.
- Coverslip in aqueous medium.

NOTE: Dehydrating in alcohol and clearing in xylene results in immediate loss of AEC from slide.

References:

- Zorina, T.D., Subbotin, V.M., Bertera, S., Alexander, A.M., Haluszczak, C., Gambrell, B., Bottino, R., Styche, A.J., Trucco, M. Recovery of the Endogenous β Cell Function in the NOD Model of Autoimmune Diabetes. *Stem Cells*, 2003, Volume 21: pages 377-388.
- Wykosky, J., Gibo, D.M., Stanton, C., Debinski, W. EphA2 as a Novel Molecular Marker and Target in Glioblastoma Multiforme. *Molecular Cancer Research*, October 2005, Volume 3: page 541. Doi: 10.1158/1541-7786.MCR-05-0056
- de Groot-Besseling, R.R.J., Ruers, T.J.M., Lamers-Elmants, I.L., Maass, C.N., de Waal, R.M.W., Westphal, J.R., Angiostatin generating capacity and anti-tumour effects of D-penicillamine and plasminogen activators. *BMC Cancer*, 2006, Volume 6: page 149. Doi: 10.1186/1471-2407-6-149
- Toyokawa, H., Nakao, A., Bailey, R.J., Nalesnik, M.A., Kaizu, T., Lemoine, J.L., Ikeda, A., Tomiyama, K., Papworth, G.D., Huang, L., Demetris, A.J., Starzl, T.E., Murase, N. Relative Contribution of Direct and Indirect Allorecognition in Developing Tolerance After Liver Transplantation. *Liver Transplantation*, March 2008, Volume 14(3): pages 346-357. doi: 10.1002/lt.21378
- Castillo, S.S., Doger, M.M., Bolken, S., Yanardag, R. Cholesterol efflux and the effect of combined treatment with niacin chromium on aorta of hyperlipidemic rats. *Molecular and Cellular Biochemistry*, 2008, Volume 308, Number 1-2: pages 151-159. doi: 10.1007/s11010-007-9623-2
- Cayli, S., Ocakli, S., Erdemir, F., Tas, U., Aslan, H., Yener, T., Karaca, Z. Development expression of p97/VCP (Valosin-containing protein) and Jab1/CSN5 in the rat testis and epididymis. *Reproductive Biology and Endocrinology*, 2011, Volume 9: page 117. doi: 10.1186/1477-7827-9-117

Bulk Reagent Ordering Information and Current Pricing at www.scytek.com

Description:	Catalog #	Volume
AEC Chromogen Concentrate	ACD015	15 ml
	ACD030	30 ml
	ACD125	125 ml
AEC Substrate Buffer	ACE500	500 ml
	ACE999	1000 ml

Storage: 2° C  8° C

 ScyTek Laboratories, Inc.
205 South 600 West
Logan, UT 84321
U.S.A.

CE IVD

EC REP

Emergo Europe
Westervoortsedijk 60
6827 AT Arnhem, The Netherlands